

Outward No. 308/2021 Date 26/2/21
 Agri-Business Incubation Centre (R-ABI)
 HeA & FW, Gol
 Dr. PDKV, Akola

"UTKARSH-PDKV" AGRIBUSINESS INCUBATION CENTRE
 Near Shahid Stambh, Opposite to Vice-Chancellor Office,
 DR.PANJABRAO DESHMUKH KRISHI VIDYAPEETH, AKOLA-444104

Quotation Notice

Subject: INVITATION OF "QUOTATION FOR ANAEROBIC WORKSTATION DISPLAY WITH O2 DISPLAY" FOR R-ABI WORKSHOP.


Sealed quotations are invited from the different contractors for anaerobic workstation display with O2 display. Please arrange to supply rates to the undersigned in prescribed following proforma. The quotation may be send to Agri-Business Incubation Centre, Dr. PDKV, Akola office on or before 4/03/2021 up to 05 pm in sealed envelope.

Particulars work

Sr. No	Item	Specification	Rate (Rs./quantity)
1	ANAEROBIC WORKSTATION DISPLAY WITH O2 DISPLAY	Anaerobic Workstation Display should be with Digital Oxygen Display to indicate absence of Oxygen in the chamber, Display should be LCD Display with touch screen Controlled, Display should have indicator for Chamber pressure, Blower Temperature, Fluorescent Light with hour meter, Blower, Temperature, Gas pressure indication to continuously monitor positive pressure inside the chamber, Positive Pressure Display should be (0.15 KPA/0.0015 bar) positive pressure of anaerobic environment for continuously maintain inside the incubation chamber, It should have Control Panel with Indicator and Low gas alarm with indicator on screen, Temperature should be ambient to 48 deg C, Temperature Uniformity should be ± 2 °C at 37°C, Temperature & Humidity should be Controlled with Display, It should have O2 sensor, Temperature sensor, Humidity sensor and Audio and visual alarm for gas cylinder pressure, It should also provide Auto excess pressure control system and Condensate controller technology, Power Supply should be 230 V \pm 15%, 50 Hz \pm 3%	

Terms and conditions:

1. The rates quoted should be inclusive of all taxes, material, any other charges if any.
2. The quotation will be accepted on all working days between 11 am to 5 pm.
3. The rates should be valid till 31/03/2021 from the date of opening of the quotations.
4. The work should be completed within time after the issue of supply/ work order.
5. Quotations duly sealed should be addressed to the undersigned and super scribed the envelope with "Quotation for anaerobic workstation display with O2 display" should reach on or before 4/03/2021 up to 5.00 pm.
6. Rates should quoted for Akola with G.S.T.


 Head & PI,

Agri-Business Incubation Centre
 Dr.PDKV, AKOLA

Copy Submitted to:

Incharge, ARIS Cell, Dr. PDKV, Akola to publish on university website