



HEAD, DEPARTMENT OF AGRONOMY
Dr. PANJABRAO DESHMUKH KRISHI VIDYAPEETH, AKOLA
P.O. Krishi Nagar, AKOLA - 444104 (Maharashtra)

NO/HAG/COART/1960/2019

Dt. 22/01/2019

QUOTATION NOTICE

To,

M/s

M/s

M/s.....

M/s.....

Subject: Quotation for Purchase of Fermenter.

Sealed quotations are invited for rate of **Fermenter** for preparation of uniform suspension of microbial inputs and packing in bottles, on or before **Dt. 04/02/2019 up to 17.00 hrs.** Quotations received after due date and time will not be considered. The quotation indicating technical details is enclosed herewith, which need to be submitting on letter head of the firm.

Sr. No.	Particulars	Specification	cost per unit
1	Fermenter	Working Capacity 200 Liter; Water Capacity 190 Ltr. Main vessel made of SS316 jacket of SS304, dish and shell of 5mm; With agitator three six bladed SS316, turbine type standard one layer for foam breaker with geared motor 3 H.P. bonfigully, premium Remi with mechanical seal silicon carbide for zero contamination with 10 mm Spurger coil for aeration with bacterial filter 0.02 micron, with glass rota meter to show airflow pressure gauge, safety valve rapture disc; sterilization can be SIP when connecting steam empty vessel sterilization and full vessel sterilization with manual valve to sterilization; CIP can be CIP when connecting cleaning system the vessels is fitted with self rotating spray balls for clearing. Broth inlet, outlet, sample outlet, Acid/alkali inlet port, Steam inlet/ water valves, pressure guage, Safety valve, Rapture disc Own make, Light glass and sight glass, all contact parts of vessel SS316, and other is SS304.	




HEAD, DEPARTMENT OF AGRONOMY
Dr. PANJABRAO DESHMUKH KRISHI VIDYAPEETH, AKOLA
P.O. Krishi Nagar, AKOLA - 444104 (Maharashtra)

		<ol style="list-style-type: none">1) Working pressure inside vessel 2kg/cm²2) Design pressure inside vessel 2.5 kg/cm²3) Test pressure 3 kg/cm²4) Working pressure inside jacket 3 kg/cm²5) Design pressure inside jacket 3.5 kg/cm²6) Test pressure 5 kg/cm²7) Design temp 130 degree.8) Auto control of temperature and variable RPM of agitator by providing four solenoid valve temp controller variable frequency drive with control panel.	
--	--	---	--

Terms and conditions for the supply of Fermenter

- 1) The rates quoted will have to be inclusive of discount rates if any and all taxes (SGST & CGST) etc.
- 2) Quotation after due date will not be considered.
- 3) If the performance of services is unsatisfactory then the decision of the university authority regarding suspend of payment will be final and binding on supply agency.
- 4) No advance to be issue for said purchasing.
- 5) Delivery must be handed over to the office within 7 days of work order.
- 6) Bill may submit in triplicate to the undersigned.
- 7) All rights are reserved with the undersigned either to accept even partly or reject totally
- 8) The rates quoted will have to be accepted for a period of six months/financial year from the date of supply order.
- 9) Authorization and Manufacturer certificate should be attached to quotation rate.
- 10) All duties, taxes and other levies payable on raw materials and components shall be included in the total price.
- 11) Sales tax in connection with the sale shall be shown separately.
- 12) The rates quoted by the bidder shall be fixed for the duration of the contract and shall not be subject to adjustment on any account.
- 13) The prices shall be quoted in Indian Rupees only.
- 14) The purchaser will evaluate and compare the quotations determined to be substantially responsive i.e. which a) Are properly signed; and b) Confirm to the terms and conditions, and Specifications. The quotations would be evaluated for all the items together/would be evaluated separately for each item.(select one of the option). Sales tax in connection with sale of goods shall not be taken into account in evaluation.
- 15) The bidder whose bid is accepted will be notified of the award of contract by the purchaser prior to expiration of the quotation validity period. The terms of the accepted offer shall be incorporated in the purchase order.
- 16) Normal commercial warranty / guarantee shall be applicable to the supplied goods.
- 17) You are requested to provide your offer latest by 17:00 hours on or before 04/02/2019.


Head
Department of Agronomy
Dr. P.D.K.V., Akola