

Quotation Notice

Subject: Permission to call quotations for purchase of Thermal imaging sensor camera under Agency trial grants available for the financial year 2020-2021 - reg

Please arrange to send quotation for **Thermal imaging sensor camera** in the old Biotechnology Centre under particulars and conditions as mentioned below. The details of the **Thermal imaging sensor camera** to be purchased are given in table on next page with form of quotation.

1. Particulars of supply to be made (Form attached).
2. The quotation should be given in the form enclosed.
3. The quotation should be in the name of **Head, Department of Agricultural Botany, Dr. PDKV, Akola mentioned "Quotation for Thermal imaging sensor camera at old Biotechnology Centre"** on envelop.
4. The quotation should reach to this office on or before **28.12.2020**.
5. The quotation will be accepted on all working days between 10.30am to 5.30pm and will be opened on the last date of receiving quotation at 4.30 pm.
6. The quoted rate should be inclusive of all taxes with delivery at Biotechnology centre, Department of Agril. Botany, Dr. PDKV, Akola.
7. The rates quoted should remain valid for a period of 6 months from the date of opening the quotation.
8. The right to accept whole or part of the quotation or reject is reserved with undersigned without assigning reason therefore.
9. The material will have to be supplied as per specifications within fifteen days from the date of issue of the order or as indicated in order.
10. In the event of failure to supply the material within the specified period the undersigned will be authorized to cancel the order for supply.
11. The quantity required is subject to variation and no reason will be assigned for it.
12. Authorized dealers in different product may send a photo copy of company rates along with quotation.
13. Samples wherever possible should be sent along with quotation.
14. **GST may be quoted on the quotation.**
15. Quotation should be in printed format of quotation.



Head
Department of Agril. Botany
Dr. P. D. K. V., Akola

FORM OF QUOTATION


Quotations for **Thermal imaging sensor camera** at old Biotechnology Centre as per the following specifications in response to enquiry of the Department of Agricultural Botany, Dr. PDKV, Akola.

SN.	Name of the item with specifications	
1.	THERMAL IMAGING SENSOR: 1 No.	
	Frame rate	: >8.0 Hz
	Spot size ratio/D:S Ratio	: 24 : 1
	Display TFT LCD	: 2.5 to 3.5 inch
	Physical shock abortion	: 2 to 3 m drop
	Accuracy	: +3% or 3 ⁰ C
	Range	: -20 to 400 °C
	Emissivity	: 0.1 to 0.09
	Resolution	: 0.1 °C
	Response time	: 150-200 millisecond
	Spectral Response	: 8 to 14µm
	Image Resolution	: 160 X 120(19,200 Pixels)
	Field of view (H X W)	: 50° X 38.6°
	Thermal imaging sensitivity	: < 70mK
	Image format	: Image with Temperature & Emissivity
	Battery	: >3 hrs of continuous use
	Memory and storage	: >5000 Images capture
	Tripod	: UNC 1/4" -20
	Holster	: Compatible
	Warranty for product	: 2 years
	Warranty for thermal imaging detector	: Minimum 10 years

Conditions in the quotation notice are accepted by me

Place:

Date:



**Signature of the Supplier
& Seal of the firm**