

## DEPARTMENT OF ENTOMOLOGY

**DR. PANJABRAO DESHMUKH KRISHI VIDYAPEETH, AKOLA**

No. HD/ENT/AUTOCLAVE/261/2024

Date: - 29/07/2024

### QUOTATION ENQUIRY NOTICE

To,

ARIS Cell (For wider publicity)  
Dr. PDKV, Akola

**Subject: Quotation called for Autoclave Vertical Type (Semi-Automatic) regarding...**


Please arrange to send the quotation of above particulars and their specification as listed below to this office before **16:00 hrs on 05/08/2024** in sealed envelope

Name of Material	Specifications	Required Quantity (Nos.)	Rate including GST with installation and transportation charges (Rs.)
Autoclave Vertical Type (Semi-Automatic)	<b>Model No. BTL-1</b> <b>Capacity: 90 Ltrs</b> <b>Inner Chamber Size: 16 dia. X 24 H inch</b> <b>Watt: 4000 W</b> Double walled construction, Inner Chamber made up of thick stainless steel (304) and the outer body is made up of thick stainless steel (304). The chamber is absolutely leaf proof and can be operated at any selected point in between 5 to 20 pounds per square inch supplied along with SS basket, Pressure gauge, release cock and safety valves (2), drain valve (Bottom and with foot arrangement, water level cutoff system)	01 (One)	

The conditions applicable for the supply of the same are mentioned below,

#### **Terms and conditions:**

1. The rates quoted should be inclusive of all levies, taxes, CGST/SGST with delivery at **Head, Department of Entomology, Dr. P. D. K. V., Akola.**
2. The rates should be **valid till 31/03/2026** from the date of opening of the quotations.
3. The right to accept or to reject whole or part of quotation is reserved with the undersigned without assigning any reason.
4. The material will have to be supplied as per the specification and make quoted in the quotation.
5. In the event of failure to supply the material within the specified time period the undersigned will have discretion to cancel the order.
6. Quotations duly sealed should be addressed to the undersigned and super scribed the envelope with "**Quotation for Autoclave Vertical Type (Semi-Automatic)**".

  
Head  
Department of Entomology  
Dr. P.D.K.V., Akola